



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Margaret T. Donnelly
Director



Jeremiah W. (Jay) Nixon
Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

CERTIFICATE OF CALIBRATION

This report shall document that the instrument described herein was examined and tested against NIST-traceable reference standards in accordance with Missouri Department of Health and Senior Services State Public Health Laboratory Breath Alcohol Program's Testing Requirements for the Recalibration of Simulators.

AGENCY: Dexter Police Department

ADDRESS: 305 Cooper St.

CITY: Dexter, MO

ZIP: 63841

SIMULATOR INFORMATION

Serial Number: SD 3328

Manufacturer: Guth

Model Number: 10-4D

CALIBRATION RESULTS

<u>Reference Temperature</u>	<u>Simulator Temperature</u>	<u>Tolerance</u>	<u>Bias</u>	<u>In Tolerance</u>	<u>Combined Uncertainty</u>
34.07	34.0	±0.1 °C	+0.07 °C	YES	±0.04 °C

This calibration was performed with NIST-Traceable Thermometer SN: 304447

This calibration was performed by: Brian Lutmer

This calibration was performed: 03/12/2013